

| | | |
|--------------------------------|---|--|
| Record Summary | 161839 | Aero Healthcare - Non-rechargeable external defibrillator electrode, adult |
| Sponsor | Aero Healthcare | |
| Therapeutic Type | Medical Device | |
| Product Category/ Class | Included Class 1 | |
| ARTG Start date | 16/05/2009 | |
| Postal Address | PO Box 1635, ARMIDALE, NSW, 2350 Australia | |
| Billing Address | PO Box 1635, ARMIDALE, NSW, 2350 Australia | |

Conditions

The automatic conditions applicable to the inclusion of all kinds of medical devices in the Register are as specified in section 41FN of the Therapeutic Goods Act 1989.

The standard conditions that are imposed under section 41FO of the Therapeutic Goods Act 1989 when kinds of medical devices are included in the Register are as set out in the following paragraphs.

For a medical device included in the Register under Chapter 4 and imported into Australia, the Sponsor must ensure that information about the Sponsor is provided in such a way as to allow the sponsor to be identified.

Each sponsor shall retain records of the distribution of all of the sponsor's medical devices included in the Register under Chapter 4. In the case of records relating to a Class AIMD medical device, Class III medical device, or Class IIb medical device that is an implantable medical device, the distribution records shall be retained for a minimum period of 10 years. In the case of records relating to any other device, the distribution records shall be retained for a minimum period of 5 years.

The sponsor of a medical device included in the Register under Chapter 4 shall keep an up to date log of information of the kind specified in Regulation 5.8.

It is a condition of inclusion in the ARTG that the sponsor of a medical device that is an AIMD, Class III or implantable Class IIb provides three consecutive annual reports to the Head of the Office of Devices, Blood and Tissues, Therapeutic Goods Administration following inclusion of the device in the ARTG. (as specified in 5.8 of the regulations) Annual reports are due on 1 October each year. Reports should be for the period 1 July to 30 June. The first report following the date of inclusion in the ARTG must be for a period of at least six months but no longer than 18 months. Subsequent reports are to be provided on 1 October for a further 2 years. The annual report must include all complaints received by the manufacturer relating to problems with the use of the device that have been received by them over the year.

Where a medical device included in the Register, contains a substance which is included in the Fourth Schedule to the Customs (Prohibited Imports) Regulations or the Eighth Schedule to the Customs (Prohibited Exports) Regulations the Sponsor shall, at the time of importation or exportation of the medical device, be in possession of a licence and a permission for importation or exportation of each consignment of the goods as required by those regulations.

A sponsor shall ensure that a medical device within their control is stored and transported in accordance with the instructions and information provided by the manufacturer.

Manufacturers

| Name | Address | Certificate number(s) |
|--------------------------------|--|-----------------------|
| HeartSine Technologies Limited | Canberra House 203 Airport Road West Belfast, Ireland, BT3 9ED Ireland | |

Products

1.Non-rechargeable external defibrillator electrode, adult

| | | | |
|---------------------|-----------------------|-----------------------|------------|
| Product Type | Single Device Product | Status | Current |
| | | Effective date | 16/05/2009 |

GMDN 47911 Non-rechargeable external defibrillator electrode, adult

Functional description Not included on record

Intended purpose An electrical conductor, with integral batteries and regulated by a dedicated external pulse generator (EPG), designed to create an electrical shock(s) and defibrillate the heart (restore normal rhythm) to treat ventricular fibrillation or pulseless ventricular tachycardia in an adult patient. It is a cartridge-type electrode, in pairs, with non-rechargeable batteries that provide the energy to produce the electrical shock(s) after its adhesive pads are placed on the skin of the patient. This is a single-use device that is replaced after a patient application or after elapse of its expiry date.

Variant information

Device Information

| | |
|----|------------------------------------|
| 10 | Single-use devices |
| 04 | Electro mechanical medical devices |

Specific Conditions

No Specific Conditions included on Record

Poison information

Poison Not recorded

Additional Product information

Record Summary